

Response ID ANON-CBS3-RZAJ-2

Submitted to Consultation: Proposed changes to the IVD medical device classifications and definitions
Submitted on 2025-03-13 12:55:04

Introduction

What is your name?

Name:

[REDACTED]

What is your email address?

Email:

[REDACTED]

What is your organisation?

Organisation:

[REDACTED]

Responding to this Consultation

Question 1(a): Do you agree with the proposals to change the Australian classification rules and principles that have an impact on approved products (as specified in the first Section of the paper), noting the changes are reflective of the regulatory scrutiny based on the associated health risks?

Respond to the question 1a:

Yes,

Question 1(b): If no, which of the proposed changes do you not agree with? Please provide your reasons.

Respond to the question 1b:

N/A

Question 1(c): Are there any other classification rules and principles, relating to the IVD medical devices, that need to be considered as part of this proposal?

Respond to the question 1c:

the activities of the UK and how they might influence the EU

Question 2(a): Do you agree with the proposals to adopt certain terminology in the Australian classification rules that have no impact on approved products (as specified in Appendix A of the paper), noting the changes are to improve clarity?

Respond to the question 2a:

yes

Question 2(b): If no, which of the proposed changes do you not agree with? Please provide your reasons.

Respond to the question 2b:

N/A

Question 2(c): Do you agree the proposed changes in Appendix A of the paper, would not result in any impact on existing ARTG entries of IVD medical devices?

Respond to the question 2c:

No, there appears to be some misalignment on classification for some products. In Australia, the original proposed cancer screening product was class III, however, with a proposed intended use change, the class would then be II. However, as it is related to Cancer, in the UK and EU, it would still be regarded as class C

Question 2(d): Are there any other classification rules, relating to the IVD medical devices, that need to be considered as part of this proposal?

Respond to the question 2d:

I do not believe so.

Question 3(a): Do you agree with the proposal to amend the Australian definitions as specified in Appendix B of the paper?

Respond to the question 3a:

yes

Question 3(b): If no, which of the proposed changes do you not agree with? Please provide your reasons.

Respond to the question 3b:

N/A

Question 3(c): Are there any other definitions, relating to the IVD medical devices, that need to be considered as part of this proposal?

Respond to the question 3c:

I do not believe so

Question 4(a): Do you agree with the proposal to apply a 6-month transition period after the EU IVDR transition timelines for the proposed Australian amendments to take effect?

Respond to the question 4a:

yes,

Question 4(b): Provide reasons for your position.

Respond to the question 4b:

as long as there is support from the TGA to educate the public on changes.

Question 5: Do you consent to your response being made publicly available on the TGA's Consultation Hub website? Please indicate your publishing preferences.

I consent to my submission being published anonymously (without my name or my organisation's name)

Question 6: If you consent to your submission being published, are there parts that you do not want published? Please specify which part(s). Please note – your contact email address and/or phone number will not be published with your submission.

Respond to the question 6:

question 2c